



Manufactured in U.S.A. by:

SCI-PHARM



UNTIL[®] Automix

Temporary Crown and Bridge Cement
in Automix Dispensing Syringe Refill
Sci-Pharm Catalog No. 55-256

TECHNICAL BULLETIN Instructions

UNTIL[®]

Automix

Temporary Crown and Bridge Cement in Automix Dispensing Syringe Refill

KIT CONTAINS:

- Cement in an Automix Dual Barrel Dispenser Refill, 6g
- Instructions

Cat. No. 55-256



INDICATIONS FOR USE:

For use as a resin-based temporary luting agent.
Contains no Eugenol. Compatible with permanent restoratives and cements.

CONTRAINDICATIONS:

This product should not be used where patients have known hypersensitivity to methacrylate monomers.
Do not apply to very sensitive teeth. For direct pulp exposure use a Calcium Hydroxide base to cap the pulp.

CAUTION: Dentin sealer contains flammable solvent.

© Copyright 2006 Scientific Pharmaceuticals, Inc.

OUTSTANDING FEATURES OF THE MATERIAL

- Ease of dispensing and no mixing required
- Good reproducibility of working time
- X-ray opacity
- Low irritation potential
- Negligible solubility
- Compatible with all commercially available permanent cements
- Low film thickness
- Well-balanced adhesive properties provide good prosthesis retention while allowing for trouble-free removal
- Excellent resistance to oral environment and mechanical strength assure good long-term performance, whenever desired

PROPERTIES OF CURED CEMENT

Water Solubility	Negligible
Film Thickness	Below 20 μ
Hardness, Barcol	40-42
Compressive Strength	69 MPa (10,000 psi)
Diametral Tensile Strength	13.8 MPa (2,000 psi)

CAUTION

GENERAL INFORMATION

Until® is a resin-based dental material specifically developed for use as a temporary luting agent. Its physical properties and handling characteristics were optimized for providing adequate working time, ease of placement and cleaning, and reliable performance, even in cases requiring extended service time.

Until® bonds to dentin and metals strongly enough to prevent marginal leakage and loss of retention, yet it allows for trouble-free removal. Good biocompatibility of this cement is expected to minimize post-operative discomfort.

Until® features a unique two-stage curing mechanism. In the first stage, the material achieves a consistency hard enough to hold the restoration in place, while allowing 60-80 seconds for easy removal of the excess.

Until® does not contain Eugenol. It was formulated for compatibility with all permanent crown and bridge cements presently in use. Its automix dispensing system facilitates handling and application.

APPLICATION

In case of direct pulp exposure, use a Calcium Hydroxide-type base to cap the pulp.

As with all synthetic cements which are hydrophobic by nature, it is imperative to disinfect and dry the preparation before cementing. The presence of moisture may result in inferior retention, possibility of microbial contamination, and post-operative sensitivity. Unless counter-indicated for the reason of proximity to the pulp, the removal of smear layer prior to cementing will further contribute to enhanced retention and greater integrity of the tooth/cement interface.

Treatment of the abutment for 30 seconds with antimicrobial solutions such as 2% Sodium Hypochlorite (NaOCl)* contributes to a lower incidence of post-operative sensitivity and prevents occasional abutment discoloration and odor due to microbial growth. On sensitive teeth and for patients with no known allergic reaction, Sodium Hypochlorite may be substituted with Erythromycin (250mg/5ml). Treatment with antimicrobial agents should be preceded by impression taking and followed by rinsing and drying.

* 2% solution of Sodium Hypochlorite may be prepared by diluting commercial 5% Hypochlorite Bleach with water in a 4:6 ratio.

INSTRUCTIONS FOR CEMENTING

1) Remove cap from the dual barrel syringe unit by turning it 90° counter-clockwise and pulling.

Before dispensing, gently push the handle until both pastes are flowing evenly from their orifices.

2) Attach the static mixer (dispenser tip) and lock it in position by turning clockwise.

NOTE: Use Sci-Pharm Static Mixers (Cat. No. 35-56) or mixers provided with Until® Automix (Cat. No. 55-26).

3) Dispense and spread evenly a desired amount of mixed cement on the internal surfaces of the restoration. Firmly set it in the mouth. Trim away excess cement, after the material starts to set (approximately 2 minutes).

4) Remove the static mixer from the dual barrel syringe unit and replace the cap.

Until® allows working time of approximately 90 seconds.

Full cure will occur within 3 minutes after the material is exposed to body temperature.

HELPFUL HINTS

- 1) If desired, bonding strength to dentin may be reduced by applying a very thin layer of releasing agent over abutment. (This procedure is especially recommended for tight-fitting temporary restorations. If the restoration is intended to be recemented, the releasing agent should also be applied inside the restoration, in order to facilitate cleaning.)
- 2) Occasionally, black stains may be observed on the abutment prepared for permanent cementation. They may be easily removed with a 2% Hydrogen Peroxide solution.

STORAGE & SHELF-LIFE

Store at temperatures not exceeding 73°F (23°C). When stored under such conditions, the material has a shelf-life of eighteen months. Refrigerate when the material is not in use (for example, overnight and on weekends). When cold, the material has a stiffer consistency. For easier handling, remove from refrigeration at least 15 minutes prior to use.



For technical information, call or write:

**Quality Management System Certified to
ISO 13485:2016**



SCIENTIFIC PHARMACEUTICALS, INC.
3221 PRODUCER WAY • POMONA, CALIFORNIA USA 91768
PHONE: (800) 634-3047 • (909) 595-9922
FAX: (909) 595-0331 • E-MAIL: scipharm@msn.com
WEBSITE: <http://www.scipharm.com>